



# Malaria Generic Pf RDT Results Guide



## NEGATIVE RESULT




## POSITIVE RESULTS



*Plasmodium falciparum*

Note: Test is positive even if the test line is faint.

## INVALID RESULTS



No Control Line

Note: Repeat the test using a new RDT if no control line appears.